



CALIBRATION RECOMMENDATIONS

Tinker & Rasor is a manufacturer of high-quality holiday detection and cathodic protection equipment. To ensure the consistent working state of our instrumentation, each instrument is hand calibrated by our expertly trained staff.

T&R recommends that each instrument be calibrated no less than once per year, beginning one year after the purchase date. To maintain warranty coverage calibration services should only be performed at our T&R factory in New Braunfels TX, USA or Authorized Service Centers.

Here is a list of instruments that are recommended for annual calibrated checks:

Holiday Detectors Models: AP*, AP-S1*, AP-S2*, APS, AP/W, M/1, M/1 Retro, M/2*, M1/AC and HD-AC

Voltmeters Models: CPV-2*, CPV-4, PRM

Other Models: AC-15, ACT, DL-1*, SR-2, SR-2V*, VC-1*, VC-2, RF-IT, CE-IT and CS-10

Each calibrated unit receives a letter and label verifying the date of calibration.

Accuracy checks performed by Tinker & Rasor factory technicians are recommended annually for other products manufactured by Tinker & Rasor that cannot be calibrated.

These instruments include:

CI-50, Quasar*, TriStar*, and pipe & cable locators.

**Indicates discontinued units*

US Tel: (833) 332-1010 International Tel: (830) 253-5621
2828 FM 758 New Braunfels, TX 78130
Web: www.tinker-rasor.com E-mail: Info@tinker-rasor.com